

# CERTIFICATE OF CONFORMITY



1. **HAZARDOUS (CLASSIFIED) LOCATION ELECTRICAL EQUIPMENT PER US REQUIREMENTS**
2. **Certificate No:** FM17US0151X
3. **Equipment:** G5 Weighing Instrument  
**(Type Reference and Name)**
4. **Name of Listing Company:** Vishay Precision Group
5. **Address of Listing Company:** 9210 Rochester Ave  
Rancho Cucamonga CA  
91730  
United States
6. The examination and test results are recorded in confidential report number:  
  
3057860 dated 12<sup>th</sup> June 2017
7. FM Approvals LLC, certifies that the equipment described has been found to comply with the following Approval standards and other documents:  
  
FM Class 3600:2011, FM Class 3611:2004, FM Class 3810:2005,  
ANSI/NEMA 250:2003, ANSI/IEC 60529:2004
8. If the sign 'X' is placed after the certificate number, it indicates that the equipment is subject to specific conditions of use specified in the schedule to this certificate.
9. This certificate relates to the design, examination and testing of the products specified herein. The FM Approvals surveillance audit program has further determined that the manufacturing processes and quality control procedures in place are satisfactory to manufacture the product as examined, tested and Approved.
10. **Equipment Ratings:**  
  
G5 Panel Mount - Non-incendive for Class I, II, III Division 2, Groups A, B, C, D, E, F, and G with Non-incendive field wiring connections to Class I, II, III, Division 2, Groups A, B, C, D, E, F, and G hazardous (classified) locations in accordance with drawing 401791;

## Certificate issued by:



J. E. Marquedant  
Manager, Electrical Systems

12 June 2017

Date

To verify the availability of the Approved product, please refer to [www.approvalguide.com](http://www.approvalguide.com)

**THIS CERTIFICATE MAY ONLY BE REPRODUCED IN ITS ENTIRETY AND WITHOUT CHANGE**

FM Approvals LLC, 1151 Boston-Providence Turnpike, Norwood, MA 02062 USA  
T: +1 (1) 781 762 4300 F: +1 (1) 781 762 9375 E-mail: [information@fmapprovals.com](mailto:information@fmapprovals.com) [www.fmapprovals.com](http://www.fmapprovals.com)

# SCHEDULE



US Certificate Of Conformity No: FM17US0151X

G5 Rail Mount - Non-incendive for Class I, Division 2, Groups A, B, C and D with Non-incendive field wiring connections to Class I, II, III, Division 2, Groups A, B, C, D, E, F, and G hazardous (classified) locations in accordance with drawing 401789.

11. The marking of the equipment shall include:  
G5 Panel Mount:

Class I, II, III, Division 2, Groups A, B, C, D, E, F, G; T4A Ta = 50°C; Type 4X, IP65

Nonincendive Field Wiring Connections to Class I, II, III, Division 2, Groups A, B, C, D, E, F, G

WHEN INSTALLED PER DWG 401791

G5 Rail Mount:

Class I, Division 2, Groups A, B, C, D; T4A Ta = 50°C

Nonincendive Field Wiring Connections to Class I, II, III, Division 2, Groups A, B, C, D, E, F, G

WHEN INSTALLED PER DWG 401789

12. **Description of Equipment:**

**General** - The G5 Weighing Instruments are single-channel or dual Channel weight indicators (G5\_PM model, panel mounted) or weight transmitters (G5\_RM model, DIN rail mounted) intended for industrial systems. The basic function is to convert the signals from strain gauge transducers to useful weight information. Transducer excitation is included as well as parameter controlled signal processing, indication of output levels, error supervision and operation of optional external equipment.

**Construction** - The G5 electronics are housed in a polymeric enclosure. The G5 Panel Mount provides an operator interface with 31 membrane keys an LCD display and must be installed in the face of a panel enclosure. The G5 Rail Mount is for installation in a suitable enclosure and has no operator interface keypad. Both types contain three circuit board assemblies, Power Board, CPU Board and Analog Board. The 110/230Vac version of the G5\_PM also has an AC Power Supply Board.

**Ratings** - The G5 PM is available in 110-240Vac or 24Vdc version, the G5 RM is only available in a 24Vdc version. The panel mount version is supplied with a Stainless Steel housing for a Type 4X, IP65 rating.

**G5\_PM\_a\_b\_c\_d. Weighing Instrument.**

a = Channels: S or D.

b = Power Supply: DC or AC.

c = Software: W, F, S or Blank.

d = Enclosure: 1, 2 or 3.

**G5\_RM\_a\_b\_c\_d. Weighing Instrument.**

a = Channels: S or D.

b = Power Supply: DC.

c = Software: W, F, S or Blank.

d = Enclosure: 1, 2 or 3.

13. **Specific Conditions of Use:**

1. *The Model G5-PM Weighing Instrument is for use indoor use only.*

2. *The Model G5-PM Weighing Instrument shall be installed in compliance with the enclosure, mounting, spacing and*

**THIS CERTIFICATE MAY ONLY BE REPRODUCED IN ITS ENTIRETY AND WITHOUT CHANGE**

FM Approvals LLC. 1151 Boston-Providence Turnpike, Norwood, MA 02062 USA

T: +1 (1) 781 762 4300 F: +1 (1) 781 762 9375 E-mail: [information@fmaprovals.com](mailto:information@fmaprovals.com) [www.fmaprovals.com](http://www.fmaprovals.com)

# **SCHEDULE**



US Certificate Of Conformity No: FM17US0151X

*segregation requirements of the ultimate application.*

*3. The USB connection shall not be used in hazardous (classified) locations.*

**14. Test and Assessment Procedure and Conditions:**

This Certificate has been issued in accordance with FM Approvals US Certification Requirements.

**15. Schedule Drawings**

A copy of the technical documentation has been kept by FM Approvals.

**16. Certificate History**

Details of the supplements to this certificate are described below:

Date	Description
12 <sup>th</sup> June 2017	Original Issue.

**THIS CERTIFICATE MAY ONLY BE REPRODUCED IN ITS ENTIRETY AND WITHOUT CHANGE**

FM Approvals LLC. 1151 Boston-Providence Turnpike, Norwood, MA 02062 USA

T: +1 (1) 781 762 4300 F: +1 (1) 781 762 9375 E-mail: [information@fmaprovals.com](mailto:information@fmaprovals.com) [www.fmaprovals.com](http://www.fmaprovals.com)